


Non-Sterile Isolation Gown



Non-woven fabric - %100 Polypropylene

One size - REGULAR

Water resistance

Disposable/Single Use

Suitable for hospital, dental and outpatient
use

Designed to protect the wearer against
airborne bacteria or virus particles



**Lightweight and comfortable
functionality & mobility**



Suitable for hospital, dental
and outpatient use

Medical
Use

!

Offers practical and
dependable
protection against
fluids, particulates
and other
contaminants.



- safety
- protection
- health


Non-Sterile Isolation Gown

High quality spunbond material, strong and durable fabric
Elastic rib cuffs and one - piece suit design
Wide range of uses in hospitals, pharmacies, dental
clinics, medical laboratories,
pet hospitals, beauty salons, research institutes.

PRODUCT PICTURES



PRODUCT PICTURES



PRODUCT DESCRIPTION

VELCRO DESIGN

User friendly velcro design.



ELASTIC CUFFS

Comfortable rib knit cuffs for flexible usage.



DRAWSTRINGS

Ease of use for each and every body shape.



PHYSICAL PROPERTIES

Features & Benefits




Non sterile isolation gown fabric provides superior protection to dust and liquid jets along. Isolation gown offers the perfect comfort and durability expected. Fabric's material allows the moisture vapor to escape through the gown. The ergonomic style pattern allows much greater freedom of movement.

BONUS ADD-ON :

Breathable fabrics and ergonomic style pattern help to reduce the risk of heat stress.

According to the fabric production plan of the mills, colours can vary. Depends on the order confirmation, we will check and inform you about the colours.



PRODUCT PICTURES




PRODUCT PICTURES



MEASUREMENTS OF GARMENT

A1	CHEST 1/2	73 cm
B	TOTAL LENGTH FROM SHOULDER	116 cm
C	TOTAL ARM LENGTH	65 cm
D	SWEEP	73 cm
E	SHOULDER TO SHOULDER	63 cm
F	TIE LENGTH	88 cm



CERTIFICATES

UNIVERSAL


UNIVERSAL
CERTIFICATION

CERTIFICATE

UNIVERSAL SERTİFİKASYON VE GÖZETİM HİZMETLERİ
TİCARET LİMİTED ŞİRKETİ,

has established and applies a medical devices Management System for the
activities below and it is certified according to UNIVERSAL CERTIFICATION procedures

" PRODUCTION AND SALES OF PROTECTIVE COVERALL, MEDICAL FACE MASKS,
BODY BAG, GOWN, NURSE UNIFORM "
EA 04

BL.02563.20 numbered report prepared as a result of the audit carried out shows that
the organization provided the requirements of

ISO 13485:2016

This certificate is valid until **04.05.2021**

Certificate Number: **20.02423**

First Date of Issue: **05.05.2020**, Recertification Date: -----, Date of Issue: **05.05.2020**


Suat KAÇMAZ
UNIVERSAL CERTIFICATION
General Manager


Dijitalizasyon kodu

The validity of this certificate is 3 years and depends on the success of the organization at the surveillance audit which will be held at least once a year.
Certificate's status can be controlled from www.universalcert.com web site or verification of QR code.
ISO 13485:2016 Standard, Edition 01-01-2016

UNIVERSAL


UNIVERSAL
CERTIFICATION

CERTIFICATE

UNIVERSAL SERTİFİKASYON VE GÖZETİM HİZMETLERİ
TİCARET LİMİTED ŞİRKETİ,

has established and applies a Good Manufacturing Practices System for the activities below and
it is certified according to UNIVERSAL CERTIFICATION procedures

" PRODUCTION AND SALES OF PROTECTIVE COVERALL, MEDICAL FACE MASKS,
BODY BAG, GOWN, NURSE UNIFORM "
EA 04

BL.02563.20 numbered report prepared as a result of the audit carried out shows that
the organization provided the requirements of

GMP – GOOD MANUFACTURE PRACTICE

This certificate is valid until **04.05.2021**

Certificate Number: **20.02423-1**

First Date of Issue: **05.05.2020**, Recertification Date: -----, Date of Issue: **05.05.2020**


Suat KAÇMAZ
UNIVERSAL CERTIFICATION
General Manager


Dijitalizasyon kodu

The validity of this certificate is 3 years and depends on the success of the organization at the surveillance audit which will be held at least once a year.
Certificate's status can be controlled from www.universalcert.com web site or verification of QR code.
ISO 9001:2015 Standard, Edition 01-01-2015

CERTIFICATES

UNIVERSALCERT.COM



ATTESTATION OF CONFORMITY

Certificate No: MDD - 109

In conformance to the European Economic Commission 93/42/EEC Medical Devices Directive on harmonisation of laws, regulations and administrative documentation of Member States on Medical Devices and European Economic Commission directive 93/68/EEC amending Medical Devices Directive dated 22 July 1993.

the products manufactured by

at the following address

EN 13795-1:2019 Surgical Clothing and Drapes - Requirements and Test Methods - Part 1: Surgical Drapes and Gowns

Brand Name:

Model: GOWN 100 Polypropylene Non-Sterile Disposable Isolation Gown

Model: GOWN 200 Polyester Non-Sterile Disposable Isolation Gown
(Standard Performance) are tested according to the following initial type tests by the manufacturer

For the assessment of conformity, the following documents were also reviewed:
Laboratory test results for Microbial Penetration (wet/dry), Bioburden,
Bursting and Tensile Strengths (wet/dry)

UNIVERSAL CERTIFICATION has evaluated production, design, intended use, risk evaluation according to safety purpose, product itself and add-on components (if exists) and product technical drawings of the surgical gowns manufactured and designed for use to prevent the transmission of infective agents between clinical staff and patients during surgical and other invasive procedures. With this certificate, it is approved that the product fulfills all essential requirements and the related rules of 93/42/EEC Medical Devices Directive (MDD) Class I are applied. The information on the packaging for the above listed products covers the necessary information stated in Annex I, §13, of the Medical Devices Directive (93/42/EEC) or Annex I, §23, of the Medical Device Regulation (EU) 2017/745. This information includes: performance level and other relevant information given in EN ISO 15223-1:2016 and EN 1041:2008+A1:2013. It is considered to be suitable to attach a CE mark, as seen below, on your products in accordance with the information given in this certificate with publishing an EU Declaration of Conformity.

This certificate is issued on 13/05/2020 and valid until 12/05/2021 with the conditions that no change has been made with the product references and no change in the production process or not suspended or withdrawn for any reason.



ISTANBUL -13/05/2020

Suat KACMAZ
UNIVERSAL CERTIFICATION
Director

This certificate will be in the absence of any changes in standard and legal terms, and with the surveillance audits to be conducted annually following the surveillance audits, updating the publication date without changing the certificate number.

EU DECLARATION OF CONFORMITY

MANUFACTURER

PRODUCT DESCRIPTION

Brand Name:

Model: GOWN 100 Polypropylene Non-Sterile Disposable Isolation Gown

Model: GOWN 200 Polyester Non-Sterile Disposable Isolation Gown

Surgical Gowns with standard performance to be used to prevent the transmission of infective agents between clinical staff and patients during surgical and other invasive procedures, classified as Medical Device (Class I)

The Manufacturer declares on his sole responsibility that the product above is, under conditions of normal use and conditions defined by the Producer / the Manufacturer, safe and meets all the necessary legal conditions and requirements. The product, a medical device that is intended for single use and solely in accordance with the Manufacturer's instructions.

The Conformity is assessed especially with the following provisions:

- European Regulation (EU) 2017/745 and 93/42/EEC Medical Devices Directive establishing technical requirements for medical devices, in effective wording
- Technical standard EN 13795-1:2019 Surgical clothing and drapes - Requirements and test methods - Part 1: Surgical drapes and gowns
- Other relevant harmonized legislation and standards
- For the assessment of conformity, the following documents were also applied to:
- Results of laboratory tests for Microbial Penetration - Dry by National Protective Testing LLC
- Results of laboratory tests for Microbial Penetration - Wet and Microbial Cleanliness, Bioburden by Ekoteks Laboratuvar ve Gözetim Hizmetleri A.Ş.
- Results of laboratory tests for Bursting and Tensile Strengths (wet/dry) by Ekoteks Laboratuvar ve Gözetim Hizmetleri A.Ş.

MARKING, LABELLING

Annex I, §13, of the Medical Devices Directive (93/42/EEC) or Annex I, §23, of the Medical Device Regulation (EU) 2017/745 specifies the information that should be specified on the packaging in which the surgical gown is supplied. The information supplied with the product considering EN ISO 15223-1:2016 and EN 1041:2008+A1:2013.

MEASURES TO ENSURE CONFORMITY

The Manufacturer declares that he has taken all necessary measures to ensure the conformity of products placed on the market with technical documentation and basic requirements for this type of product.

Gökçen Manager
13/05/2020



Test
TS EN ISO IEC 17025
AB-0716-T

AB-0716-T
TURT200061245
05-20

Page 1 of 9

TEST REPORT

REPORT NUMBER : TURT200061245

APPLICANT NAME :

ADDRESS :

Attention :

SAMPLE DESCRIPTION : One sample of Blue non sterile medical gown
DATE IN : 4 May ,2020 (13:59:00)
RESUBMIT DATE : 11 May ,2020
DATE OUT : 12 May ,2020
BUYER'S NAME :
BUYER'S REGION : EUROPE
MODEL/ STYLE NO :
FIBER COMPOSITION : Claimed to be 100% Polypropylen

Burcu KESKİNSOY KUNDAKÇI
Customer Care Executive

Zeynep AKIN
Chemical Laboratory Manager

Intertek Test Hizmetleri A.Ş.
Merkez Mahallesi Sanayi Cad. No.23 Altındag Plaza Yenibosna-34197 /İSTANBUL
Phone : +90 212 496 46 46 Fax: +90 212 452 80 55
e-mail : intertekg.turkiye@intertek.com
<http://www.intertek-turkey.com>



200061245



TEST REPORT

REPORT : TURT200061245

Page 2 of 9

TEST	SAMPLE
Determination of Alkylphenolethoxylates (APEO) for Textile	1
Determination of Cadmium Content	NR
Determination of Certain Aromatic Amines Derived from Azo Colorants	P
Determination of Free and Hydrolised Formaldehyde Test (Water extraction method)	P
Determination of Polychlorophenols	NR
Polycyclic Aromatic Hydrocarbons (PAHs) Analysis	P
Total Lead in Non Metal Products	P

P = MEETS BUYER' S REQUIREMENT / F = DOES NOT MEET BUYER' S REQUIREMENT / NR = NO REQUIREMENT / SC=STILL CONTINUES / X=NOT PERFORMED / NA = NOT APPLICABLE / LS = LACK OF SAMPLE / NC = NO COMMENT / I = INCONCLUSIVE / # = SEE RESULT / NF = NEEDS FURTHER TESTING / A = ABSENT / M = MARGINAL ACCEPT / SD = SEE DETAILS ENCLOSED / FS: FURTHER STEPS

This report (including any enclosures and attachments) are prepared for the exclusive use of the Customer(s) named in the report and solely for the purpose for which it is provided and on the basis of instructions and information and/or materials supplied by Intertek's Customer. The test results relate only to the specific items tested and are not intended to be a recommendation for any particular course of action. Customer is responsible for acting as it sees fit on the basis of such results. Unless Intertek provide express prior written consent, no part of this report should be reproduced, distributed or communicated to any third party, nor could it be used for PR activities. Intertek do not accept any liability if this report is used for an alternative purpose from which it is intended, nor do Intertek owe any duty of care to any third party in respect of this report. Except where explicitly agreed in writing, all work and services performed is governed by Intertek Standard Terms and Conditions of Service which is available on request or can be obtained at <http://www.intertek.com/terms>. Testing reports without signature are not valid. The sample has been provided by the customer and the results apply to the sample as received. Sample information is supplied by the customer. Unless otherwise requested, this laboratory applies shared risk decision rule. Tests marked (*) in this test report are not included in the TÜRKAK accreditation schedule for this laboratory.

Intertek Test Hizmetleri A.Ş.
Merkez Mahallesi Sanayi Cad. No.23 Altındag Plaza Yenibosna-34197 /İSTANBUL
Phone : +90 212 496 46 46 Fax: +90 212 452 80 55
e-mail : intertekg.turkiye@intertek.com
<http://www.intertek-turkey.com>



200061245

RESULTS
REPORT :TURT200061245

Page 3 of 9
12 May ,2020

Test Method	Results	Requirements
-------------	---------	--------------

Determination of Alkylphenoethoxylates (APEO) for Textile

BS EN ISO 18254-1:2016

Determination of APEO by Liquid Chromatography-Mass Spectrometry (LC-MS-MS) Analysis

Blue interlining, white sleeve part, white trim

Alkylphenols

Nonylphenol (NP)	Not Detected	No Requirement
Octylphenol (OP)	Not Detected	

Alkylphenol Ethoxylates

Nonylphenoethoxylates (NPEO)	Not Detected	No Requirement
Octylphenoethoxylates (OPEO)	Not Detected	

ppm = mg/kg
Detection Limit = 50 ppm
ND = not detected

Determination of Alkylphenoethoxylates (APEO) for Textile

BS EN ISO 18254-1:2016

Determination of APEO by Liquid Chromatography-Mass Spectrometry (LC-MS-MS) Analysis

White loop

Alkylphenols

Nonylphenol (NP)	Not Detected	No Requirement
Octylphenol (OP)	Not Detected	

Alkylphenol Ethoxylates

Nonylphenoethoxylates (NPEO)	Not Detected	No Requirement
Octylphenoethoxylates (OPEO)	Not Detected	

ppm = mg/kg
Detection Limit = 50 ppm
ND = not detected

RESULTS
REPORT :TURT200061245

Page 4 of 9
12 May ,2020

Test Method	Results	Requirements
-------------	---------	--------------

Determination of Cadmium Content

BS EN 1122 : 2001 (Method B)

Determination by microwave digestion and ICP -OES

White velcro	<5 ppm	100 ppm (0.01%)
--------------	--------	-----------------

ppm (part per million) =mg / kg Detection Limit =5 ppm < =Less Than
% =Percentage based on dry weight of sample
REMARK As per Cadmium Content Requirement in Annex XVII item 23 of the REACH Regulation (EC)
No:1907/2006 (Formerly Known as Directive 91/338/EEC), Acid Digestion Method was used Total Cadmium Content was determined by ICP-OES

RESULTS
REPORT :TURT200061245

Page 5 of 9
12 May ,2020

Test Method	Results	Requirements
-------------	---------	--------------

Determination of Certain Aromatic Amines Derived from Azo Colorants

EN ISO 14362-1:2017

By Gas Chromatographic - Mass Spectrometric (GC-MS) And High Performance Liquid Chromatographic (HPLC) Analysis.

1-Blue interlining (with extraction)

<30 ppm

FORBIDDEN AMINE	CAS NO	RESULTS
4-AMINOBIIPHENYL	92-67-1	N
BENZIDINE	92-87-5	N
CHLORO-O-4-CHLOR-O-TOLUIDINE	95-69-2	N
2-NAPHTHYLAMINE	91-59-8	N
*O-AMINOAZOTOLUENE	97-56-3	N
*2-AMINO-4-NITROTOLUENE	99-55-8	N
P-CHLOROANILINE	106-47-8	N
2,4-DIAMINOANISOLE	615-05-4	N
4,4'-DIAMINOBIIPHENYLMETHANE	101-77-9	N
3,3'-DICHLOROBENZIDINE	91-94-1	N
3,3'-DIMETHOXYBENZIDINE	119-90-4	N
3,3'-DIMETHYLBENZIDINE	119-93-7	N
3,3'-DIMETHYL-4,4' DIAMINOBIIPHENYLMETHANE	838-88-0	N
P-CRESIDINE	120-71-8	N
4,4'-METHYLENE-BIS-(2 CHLOROANILINE)	101-14-4	N
4,4'-OXYDIANILINE	101-80-4	N
4,4'-THIODIANILINE	139-65-1	N
O-TOLUIDINE	95-53-4	N
2,4-TOLUENEDIAMINE	95-80-7	N
2,4,5-TRIMETHYLANILINE	137-17-7	N
O-ANISIDINE	90-04-0	N
**P-AMINOAZOBENZENE	60-09-3	N
2,4 XYLIDINE	95-68-1	N
2,6 XYLIDINE	87-62-7	N

Note:
1)The amines o-amino-azotoluene and 2-amino-4-nitrotoluene are detected by its splitted product o-toluidine and 2,4- toluediamine.
2)Azo colorants that are able to form 4-aminoazobenzene, generate under the condition of this method aniline and 1,4- phenylenediamine . The presence of these colorants can not be reliably ascertained without additional information, e.g. chemical structure of the colorant used.
3)According to ISO 14362-1:2017, separate test is suggested to ascertain the compliance for result of mixed test in the range between 5 ppm and 30 ppm.
4)Azo colourants Content Requirement In Annex XVII Item 43 Of The REACH Regulation (EC) NO. 1907/2006 & Amendment No. 552/2009 and 126/2013 (Formerly Known As Directive 2002/61/EC
5) According to the official method ISO 14362-1:2017, if 4-Aminodiphenyl or 2-Naphthylamine or 2,4-Diaminoanisole is found exceeding requirement, the use of forbidden Azo colourants cannot be ascertained without additional information e.g. The chemical structure of the colourant used.

ppm : part per million (mg/kg) Detection Limit: 5 ppm < = Less Than N: Not Detected NC: No Comment

RESULTS
REPORT :TURT200061245

Page 6 of 9
12 May ,2020

Test Method	Results	Requirements
-------------	---------	--------------

Determination of Free and Hydrolised Formaldehyde Test (Water extraction method)

BS EN ISO 14184 - 1 :2011 Free and Hydrolized Formaldehyde by UV-VIS Analysis

Blue interlining, white sleeve	Not Detected	No Requirement
part, white trim		
White loop	Not Detected	

ppm (part per million) =mg / kg
Detection Limit =5 ppm
< =Less Than
Note :Sample was received unsealed

RESULTS
REPORT : TUR200061245

Page 7 of 9
12 May ,2020

Test Method	Results	Requirements
-------------	---------	--------------

Determination of Polychlorophenols

BVL B 82.02-8:2001-06

Blue interlining, white sleeve part, white trim	Result	Requirement
Pentachlorophenol (PCP)	Not Detected	0.5 ppm
Tetrachlorophenol (TeCP)		0.5 ppm
2,3,4,5 Tetrachlorophenol	Not Detected	
2,3,4,6 Tetrachlorophenol	Not Detected	
2,3,5,6 Tetrachlorophenol	Not Detected	
Trichlorophenol (TriCP)		2 ppm
2,3,4 Trichlorophenol	Not Detected	
2,3,5 Trichlorophenol	Not Detected	
2,3,6 Trichlorophenol	Not Detected	
2,4,5 Trichlorophenol	Not Detected	
2,4,6 Trichlorophenol	Not Detected	
3,4,5 Trichlorophenol	Not Detected	
Dichlorophenol (DCP)		3 ppm
2,3- Dichlorophenol	Not Detected	
2,4- Dichlorophenol	Not Detected	
2,5- Dichlorophenol	Not Detected	
2,6- Dichlorophenol	Not Detected	
3,4- Dichlorophenol	Not Detected	
3,5- Dichlorophenol	Not Detected	
Monochlorophenols (MCP)		
2-Chlorophenol	Not Detected	
3-Chlorophenol	Not Detected	
4-Chlorophenol	Not Detected	

ppm (part per million) =mg / kg Detection Limit = PCP, TeCP : 0.05 ppm (lower limit) or 0.5 ppm ; OPP : 0.5 ppm
< = Less Than ND = Not Detected
Remark : PCP Content Requirement in Annex XVII Item 22 of the Reach Regulation (EC) No : 1907/2006

RESULTS
REPORT : TUR200061245

Page 8 of 9
12 May ,2020

Test Method	Results	Requirements
-------------	---------	--------------

Determination of Polychlorophenols

BVL B 82.02-8:2001-06

White loop	Result	Requirement
Pentachlorophenol (PCP)	Not Detected	0.5 ppm
Tetrachlorophenol (TeCP)		0.5 ppm
2,3,4,5 Tetrachlorophenol	Not Detected	
2,3,4,6 Tetrachlorophenol	Not Detected	
2,3,5,6 Tetrachlorophenol	Not Detected	
Trichlorophenol (TriCP)		2 ppm
2,3,4 Trichlorophenol	Not Detected	
2,3,5 Trichlorophenol	Not Detected	
2,3,6 Trichlorophenol	Not Detected	
2,4,5 Trichlorophenol	Not Detected	
2,4,6 Trichlorophenol	Not Detected	
3,4,5 Trichlorophenol	Not Detected	
Dichlorophenol (DCP)		3 ppm
2,3- Dichlorophenol	Not Detected	
2,4- Dichlorophenol	Not Detected	
2,5- Dichlorophenol	Not Detected	
2,6- Dichlorophenol	Not Detected	
3,4- Dichlorophenol	Not Detected	
3,5- Dichlorophenol	Not Detected	
Monochlorophenols (MCP)		
2-Chlorophenol	Not Detected	
3-Chlorophenol	Not Detected	
4-Chlorophenol	Not Detected	

ppm (part per million) =mg / kg Detection Limit = PCP, TeCP : 0.05 ppm (lower limit) or 0.5 ppm ; OPP : 0.5 ppm
< = Less Than ND = Not Detected
Remark : PCP Content Requirement in Annex XVII Item 22 of the Reach Regulation (EC) No : 1907/2006

RESULTS

REPORT :TURT200061245

Page 9 of 9

12 May ,2020

Test Method	Results	Requirements
-------------	---------	--------------

Polycyclic Aromatic Hydrocarbons (PAHs) Analysis

INTERTEK IHTM AL.2.032 based on A/PS GS 2014:01

Blue interlining	Result	Requirement
BENZO(A)PYRENE	Not Detected	1 ppm
BENZO(E)PYRENE	Not Detected	1 ppm
BENZO(A)ANTHRACENE	Not Detected	1 ppm
BENZO(B)FLUORANTHENE	Not Detected	1 ppm
BENZO(J)FLUORANTHENE	Not Detected	1 ppm
BENZO(K)FLUORANTHENE	Not Detected	1 ppm
CHRYSENE	Not Detected	1 ppm
DIBENZO(A,H)ANTHRACENE	Not Detected	1 ppm

ppm (part per million) =mg / kg
Detection Limit = 0.1 ppm

Total Lead in Non Metal Products

CPSC-CH-E1002-08.3 Method followed by ICP-OES

White velcro Not Detected 100 ppm (0.01%)

ppm (part per million) = mg / kg
< = Less Than
Detection Limit = 20 ppm

END OF TEST REPORT





**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**
Esenyurt Firuzköy Bulvarı No:29 34325 Avcılar
İstanbul/ TÜRKİYE

20014347-
ing-Add

EKOTEKS

TEST REPORT
DENEY RAPORU

05-20

Customer name: UNIVERSAL SERTİFİKASYON VE GÖZETİM HİZMETLERİ TİCARET LTD.ŞTİ.
Address: 15 Temmuz Mah. Gülbahar Cad. No:96 Bağcılar/İSTANBUL
Buyer name:
Contact Person:
Order No: -
Article No: -
Name and identity of test item: Sample 1: Dark blue overall.
Sample 2: Light blue overall.
The date of receipt of test item: 06.05.2020
Re-submitted/re-confirmation date: -
Date of test: 06.05.2020-13.05.2020
Remarks: -
Sampling: The results given in this report belong to the received sample by vendor.
End-Use: -
Care Label: Not specified.
Number of pages of the report: 9

Seal

Date
13.05.2020

Head of Testing Laboratory
Sevim A. RAZAK
13.05.2020

This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

20014347-
ing-Add

05-20

REQUIRED TESTS	RESULT	COMMENTS
MICROBIOLOGICAL TEST/Sample 1&2		
Microbial Cleanliness (Bioburden) ⁽¹⁾	P	
Wet-Bacterial Penetration ⁽¹⁾	P	
PHYSICAL PROPERTIES TESTS /Sample 1&2		
Tensile Strength / Dry	P	
Tensile Strength / Wet	P	
Bursting Strength / Dry	P	
Bursting Strength / Wet	P	
P: Pass F: Fail R: Refer to retailer technologist. ⁽¹⁾ This report was reissued to add this test result. Test results were evaluated according to EN 13795-1:2016 Standard Performance Properties Critical Sample Group limit values (Table 1)		

REMARK: Original samples are kept for 3 months and all technical records are kept for 5 years unless otherwise specified. If requested, measurement uncertainty will be reported. But unless otherwise specified, measurement uncertainty is not considered while stating compliance with specification or limit values. The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximately 95 %. Tests marked (*) in this report are not included in the accreditation schedule.



Sample 2

Sample 1

This report shall not be reproduced other than in full except with the permission of the laboratory.
Testing reports without signature and seal are not valid.

TEST RESULTS

MICROBIAL CLEANLINESS (Bioburden)

Test Metod: Ref: EN ISO 11737-1:2018 (*)

The sample is put in extraciton liquid after shaking well, inoculated on the agar.
After incubation at 30 ± 1 °C for 72 hours, growth microorganisms are counted on the agar.

Sample 1		
	<u>RESULTS</u>	<u>REQUIREMENT</u>
Microbial cleanliness (cfu/g)	170 cfu/100 cm ²	≤300 cfu/100 cm ²

*cfu= Colony forming unit.

MICROBIAL CLEANLINESS (Bioburden)

Test Metod: Ref: EN ISO 11737-1:2018 (*)

The sample is put in extraciton liquid after shaking well, inoculated on the agar.
After incubation at 30 ± 1 °C for 72 hours, growth microorganisms are counted on the agar.

Sample 2		
	<u>RESULTS</u>	<u>REQUIREMENT</u>
Microbial cleanliness (cfu/g)	150 cfu/100 cm ²	≤300 cfu/100 cm ²

*cfu= Colony forming unit.

TEST RESULTS

Sample 1

Test Method: BS EN 22610: 2006 (Surgical drapes, garments and fresh air clothes used as medical devices for patients, hospital staff and equipment - Test method for determination of resistance to wet bacterial permeability) (*)

A test sample is placed on the agar plate on a rotating disc. Bacteria carrier material and coating film are placed on the test sample and all parts are fixed on the disk. A finger is placed on the test sample to apply a certain force ($3N \pm 0.02$). The finger moves on the test sample over the entire surface of the agar within 15 minutes. 5 studies are carried out for 15 minutes. 6. The study is repeated by inverting the sample.

Sample amount:	5 pieces 25x25cm ²
Carrier Material:	30 µm thin, 25x25cm ² Polyurethane Film
Coating Material:	25x25cm ² HDPE Film
Microorganism:	Staphylococcus aureus ATCC 29213
Bacterial Concentration (kob / ml):	1-4x10 ⁴ kob / ml
Incubation Conditions:	(36 ± 1) °C 48 hours

RESULTS			
Number of Populating Bacteria (cfu)		Penetration Rate	
X1	100	RCUM1	0,1
X2	0	RCUM2	0,1
X3	200	RCUM3	0,4
X4	0	RCUM4	0,4
X5	150	RCUM5	0,6
Z	260		
T	710		
X1 X5: Number of colonies growing in 5 parallel petri in the same sample			
Z: number of colonies growing in the sixth petri dish			
T: X1 + X2 + X3 + X4 + X5 + Z			
RCUM1 = X1/T			
RCUM2 = (X2 + X1)/T			
RCUM3 = (X3 + X2 + X1)/T			
RCUM4 = (X4 + X3 + X2 + X1)/T			
RCUM5 = (X5 + X4 + X3 + X2 + X1)/T			
BARRIER INDEX (IB)			
	Result	Expected value (*)	
IB	4,4	≥2,8	
IB = 6 - (CUM1 + CUM2 + CUM3 + CUM4 + CUM5)			
* EN 13795-1:2019 Surgical gowns and drapes - Requirements and test methods are evaluated according to Table-1.			

20014347- ing-Add
05-20

TEST RESULTS

Sample 2

Test Method: BS EN 22610: 2006 (Surgical drapes, garments and fresh air clothes used as medical devices for patients, hospital staff and equipment - Test method for determination of resistance to wet bacterial permeability) (*)

A test sample is placed on the agar plate on a rotating disc. Bacteria carrier material and coating film are placed on the test sample and all parts are fixed on the disk. A finger is placed on the test sample to apply a certain force (3N ± 0.02). The finger moves on the test sample over the entire surface of the agar within 15 minutes. 5 studies are carried out for 15 minutes. 6. The study is repeated by inverting the sample.

Sample amount:	5 pieces 25x25cm2
Carrier Material:	30 µm thin, 25x25cm2 Polyurethane Film
Coating Material:	25x25cm2 HDPE Film
Microorganism:	Staphylococcus aureus ATCC 29213
Bacterial Concentration (kob / ml):	1-4x104 kob / ml
Incubation Conditions:	(36 ± 1) °C 48 hours

RESULTS			
Number of Populating Bacteria (cfu)		Penetration Rate	
X1	100	RCUM1	0,1
X2	0	RCUM2	0,1
X3	200	RCUM3	0,3
X4	250	RCUM4	0,5
X5	300	RCUM5	0,7
Z	290		
T	1140		
X1 X5: Number of colonies growing in 5 parallel petri in the same sample			
Z: number of colonies growing in the sixth petri dish			
T: X1 + X2 + X3 + X4 + X5 + Z			
RCUM1 = X1/T			
RCUM2 = (X2 + X1)/T			
RCUM3 = (X3 + X2 + X1)/T			
RCUM4 = (X4 + X3 + X2 + X1)/T			
RCUM5 = (X5 + X4 + X3 + X2 + X1)/T			
BARRIER INDEX (IB)			
	Result	Expected value (*)	
IB	4,3	≥2,8	
IB = 6 - (CUM1 + CUM2 + CUM3 + CUM4 + CUM5)			
* EN 13795-1:2019 Surgical gowns and drapes - Requirements and test methods are evaluated according to Table-1.			

Genl/G3203

20014347- ing-Add
05-20

TEST RESULTS

TEST METHOD : EN 13795-1:2019

SURGICAL CLOTHING AND DROPEs -REQUIREMENTS AND TEST METHODS

ANNEX 1: SURGICAL CLOTHING AND DROPEs (*);

TENSILE STRENGTH; EN 29073-3:1996 (*)

Instron 5969 (Load: 50 kN), Strip Method.
Speed: 100 mm/min±10,Gauge length 200 mm.
Pre-load was not applied. Without wetting samples.
The average results are given for weft and warp direction of five samples
Performed in the conditioned room (20±2°C-65%±4).

Dry ;

Sample 1	RESULT
Weft	44.0 N
Warp	46.0 N

REQUIREMENT

≥ 20N (Dry)
≥ 20N (Dry)

TENSILE STRENGTH; EN 29073-3:1996 (*)

Instron 5969 (Load: 50 kN), Strip Method.
Speed: 100 mm/min±10,Gauge length 200 mm.
Pre-load was not applied. With wetting samples.
The average results are given for weft and warp direction of five samples
Performed in the conditioned room (20±2°C-65%±4).

Wet ;

Sample 1	RESULT
Weft	41.3 N
Warp	48.8 N

REQUIREMENT

≥ 20N(Wet)
≥ 20N(Wet)

BURSTING STRENGTH;; ISO 13938-1:1999

SDL ATLAS M229 tester. Test area: 30.5 mm diameter
Rate of increase in volume; 29 cm³/min.
The average results are given of five samples.
Performed in the conditioned room (20±2°C-65%±4).

	RESULT
Dry ;	135.1 kPa

REQUIREMENT

≥ 40 kPa (Dry)

Height at Burst*	7.9 mm
------------------	--------

Genl/G3203

20014347- ing-Add
05-20

TEST RESULTS

TEST METHOD : EN 13795-1:2019

SURGICAL CLOTHING AND DROPEs -REQUIREMENTS AND TEST METHODS

ANNEX 1: SURGICAL CLOTHING AND DROPEs (*);

BURSTING STRENGTH;; ISO 13938-1:1999

SDL ATLAS M229 tester. Test area: 30.5 mm diameter
Rate of increase in volume; 45.2 cm³/min.
The average results are given of five samples.
Performed in the conditioned room (20±2 °C-65%±4).

Sample 1

	RESULT
Wet ;	143.3 kPa
Height at Burst*	8.4 mm

REQUIREMENT
≥ 40 kPa (Wet)

Genl/63-2/03

20014347- ing-Add
05-20

TEST RESULTS

TEST METHOD : EN 13795-1:2019

SURGICAL CLOTHING AND DROPEs -REQUIREMENTS AND TEST METHODS

ANNEX 1: SURGICAL CLOTHING AND DROPEs (*);

TENSILE STRENGTH; EN 29073-3:1996 (*)

Instron 5969 (Load: 50 kN), Strip Method.
Speed: 100 mm/min±10,Gauge length 200 mm.
Pre-load was not applied. Without wetting samples.
The average results are given for weft and warp direction of five samples
Performed in the conditioned room (20±2 °C-65%±4).

Dry ;

Sample 2	RESULT
Weft	79.4N
Warp	78.2N

REQUIREMENT
≥ 20N (Dry)
≥ 20N (Dry)

TENSILE STRENGTH; EN 29073-3:1996 (*)

Instron 5969 (Load: 50 kN), Strip Method.
Speed: 100 mm/min±10,Gauge length 200 mm.
Pre-load was not applied. With wetting samples.
The average results are given for weft and warp direction of five samples
Performed in the conditioned room (20±2 °C-65%±4).

Wet ;

Sample 2	RESULT
Weft	85.6 N
Warp	87.0 N

REQUIREMENT
≥ 20N (Wet)
≥ 20N (Wet)

BURSTING STRENGTH;; ISO 13938-1:1999

SDL ATLAS M229 tester. Test area: 30.5 mm diameter
Rate of increase in volume; 29 cm³/min.
The average results are given of five samples.
Performed in the conditioned room (20±2 °C-65%±4).

Sample 2

	RESULT
Dry ;	136.7 kPa
Height at Burst*	14.2mm

REQUIREMENT
≥ 40 kPa (Dry)

Genl/63-2/03

20014347- ing-Add
05-20

TEST RESULTS

TEST METHOD : EN 13795-1:2019

SURGICAL CLOTHING AND DROPEŞ -REQUIREMENTS AND TEST METHODS

ANNEX 1: SURGICAL CLOTHING AND DROPEŞ (*);

BURSTING STRENGTH;; ISO 13938-1:1999

SDL ATLAS M229 tester. Test area; 30.5 mm diameter
Rate of increase in volume; 45.2 cm³/min.
The average results are given of five samples.
Performed in the conditioned room (20±2 °C-65%±4).
Sample 2

	RESULT
Wet ;	170.3 kPa
Height at Burst*	21.9 mm

REQUIREMENT
≥ 40 kPa (Wet)